S Series™ Transducer Holder


Assembly Instructions

Introduction

Attach the transducer holder to the S Series[™] ultrasound system before attaching the system to the S Series Stand or other supporting device.

See the S Series Stand User Guide or VHM™ Wall Mount Installation Instructions for S Series™ Ultrasound System for information on cleaning and disinfecting and on setting up the ultrasound system to use accessories and peripherals.

Contents



Attach transducer holder to system

- Place the transducer holder on the back of the system, with the cup facing toward the front.
 The two flat pegs above the transducer connector pass through the large holes on the holder.
- 2 Insert the Phillips bit into the magnetic screwdriver.
- 3 Using the magnetic screwdriver, tighten the four 8mm Phillips-head screws. See Figure 1.

To remove the transducer holder, loosen the four screws on the back of the transducer holder by using a Phillips screwdriver. See Figure 1.




Figure 1 Tighten transducer holder.



S Series, SonoSite, and the SonoSite logo are registered trademarks or trademarks of SonoSite, Inc.

 $Non-Sono Site\ product\ names\ may\ be\ trademarks\ or\ registered\ trademarks\ of\ their\ respective\ owners.$

This SonoSite product(s) referenced in this document may be covered by one or more of the following U.S. patents: 5722412, 5817024, 5893363, 6135961, 6203498, 6364839, 6371918, 6383139, 6416475, 6447451, 6471651, 6569101, 6648826, 6575908, 6604630, 6817982, 6835177, 6962566, 7169108, 7449640, D456509, D461895, D509900, D538432, D544962, D558351, D559390, and by the following counterpart foreign patents: AU727381, AU730822, CA2373065, CN98106133.8, CN98108973.9, DE60021552.0, DE60029777.2, DE60034670.6, DE69730563.5, DE6980539.6, DE69831698.3, FR0815793, FR0875203, FR0881492, FR1175713, FR1180970, GB0815793, GB0875203, GB0881492, GB1175713, GB1180970, GB1180971, IT0815793, IT0881492, IT1175713, KR528102, KR532359, NO326202, NZ542968, RCD000897368-0001, SP0815793, SP0881492. Patents pending.

Caution:

Federal (United States) law restricts this device to sale by or on the order of a physician.

SonoSite, Inc.

21919 30th Drive SE Bothell, WA 98021 USA T: 1-888-482-9449 or

1-425-951-1200 F: 1-425-951-1201 SonoSite Ltd

Alexander House 40A Wilbury Way Hitchin, Herts SG4 OAP UK

T: +44-1462-444800 F: +44-1462-444801 03/2009 Copyright 2009 by SonoSite, Inc. All rights reserved.



P12237-01

