

**Manufacturer's Declaration of Conformity**  
**Full Quality Assurance Procedure**

According To Australian Therapeutic Goods (Medical Devices) Regulations 2002



This is a declaration made in accordance with the requirements of Clause 6.6 of Schedule 3 of the Australian *Therapeutic Goods (Medical Devices) Regulations 2002* relating to the devices stated in following Schedule.

**Manufacturer's Name & Address:**

FUJIFILM SonoSite, Inc.  
21919 30<sup>th</sup> Drive SE  
Bothell, Washington 98021-3904 USA

**Australasia Sponsor's Name & Address:**

FUJIFILM SonoSite Australasia Pty, Ltd.  
Suite 9, 13A Narabang Way  
Belrose, NSW 2085, Australia

**Declares that the product(s):**

Item	Classification	Part Number(s)	GMDN Code
X-Porte Control Panel Assembly	Rule 1	P14027	40761,40971
Display, Clinical Monitor, LED Backlight	Rule 1	P14627	40761,36612
TTC Assembly, X-Porte	Rule 1	P14644	40761
X-Porte Dock with DVR	Rule 1	P15164	40761
Kit, USB Bar Code Scanner, JadaK	Rule 1	P14166	37035
Assembly, Cable, ECG, EU, Lead-II	Rule 1	P15171	35562
PowerPark Dock/Stand Modules	Rule 1	P12834, P12822	40761
External AC Power Supply, X-Porte	Rule 1	P14521	40761,36545

**Scope and Compliance:**

Each medical device part number listed above complies with the Therapeutic Goods (Medical Devices) Regulations 2002 and are Class I medical devices in accordance with Schedule 2, Part 1, Rule 2.1 and Rule 4.1, as demonstrated in accordance Schedule 3, Part 6, Clause 6.6 and Schedule 3, Part 1. In addition, each kind of medical device to which the Full Quality Assurance Procedures have been applied complies with the applicable provisions of the essential principles, the classification rules, at each stage, from the design of the device until its final inspection before being supplied.

**Conformity Assessment Certificate:**

CE 02429 Full Quality Assurance Certificate first issued 13 September 1999 by BSI in respect of the design, development and manufacture of portable diagnostic ultrasound equipment, transducers and other associated accessories.

**Standards Applied:**

- EN IEC 60601-1: 2006, 3<sup>rd</sup> Edition
- EN IEC 60601-2-37: 2008
- EN IEC 60601-1-2: 2007
- EN ISO 14971: 2012

**Quality Management and Quality Assurance:**

- ISO 9001:2008, Quality management systems -- Requirements International Organization for Standardization
- ISO 13485:2003, Medical devices – Quality management systems – Requirements for regulatory purposes
- EN ISO 13485:2012, Medical devices – Quality management systems – Requirements for regulatory purposes

September 5, 2013  
Bothell, Washington USA

Ken Hansen  
Director, Regulatory Affairs and Quality Systems