



Certificate of Compliance

Certificate: 2463874

Master Contract: 201161

Project: 2463874

Date Issued: March 29, 2012

Issued to: SonoSite, Inc.

21919 30th Dr SE
Bothell, WA 98021-3904
USA
Attention: Jean Bishop

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Mehrdad Sadeghieh

Issued by: Mehrdad Sadeghieh, P. Eng.

PRODUCTS

CLASS 8780 01 - MEDICAL ELECTRICAL EQUIPMENT/SYSTEMS

CLASS 8780 81 - MEDICAL ELECTRICAL EQUIPMENT/SYSTEMS - Certified to US Standards

Diagnostic Ultrasound System Model M-Turbo Ultrasound System, rated 100-240 V, 50-60 Hz, 2.0- 1.0 A or Battery Operated.

1. Type of protection against electric shock: Class I (when powered from power supply or part of the Mobile Docking System); and Internally powered (Ultrasound System w/o power supply).
2. System includes optional M-Turbo MiniDock, Mobile Docking System, and accessories.
3. Degree of protection against electric shock: Type BF transducers / Defibrillation-proof Type CF ECG leads
4. Degree of protection against ingress of water: IPX7 for ultrasound transducers/probes only, IPX8 footswitch.
5. Degree of safety of application in the presence of a flammable anaesthetic mixture with air or with oxygen or nitrous oxide: Equipment not suitable for use in the presence of a flammable anaesthetic mixture with air or with oxygen or nitrous oxide.



Certificate: 2463874

Master Contract: 201161

Project: 2463874

Date Issued: March 29, 2012

6. Mode of operation: Continuous to 35C. Non-Continuous >35 to 40C (30 min on/30 min off).
7. Environmental Conditions: Normal: 10-40°C, 15-95% (non-condensing) rH, 700-1060hPa

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 60601-1-08 Medical Electrical Equipment - Part 1: General Requirements for basic safety and essential performance

CAN/CSA C22.2 60601-1-6-07 Medical Electrical Equipment part 1-6: General requirements for basic safety and essential performance - Collateral Standard: Usability adopted IEC 60601-1-6 (06) Ed. 2

CAN/CSA C22.2 60601-2-37:08 Medical Electrical Equipment part 2-37: Particular requirements for the basic safety and performance of ultrasonic medical diagnostic and monitoring equipment, adopted IEC 60601-2-37(07) Ed 2.0

ANSI/AAMI ES60601-1:2005 (IEC 60601-1:2005, MOD) Medical electrical equipment, Part 1: General requirements for basic safety and essential performance