



CSA INTERNATIONAL

Certificate of Compliance

Certificate: 1338884 (LR 115987)

Master Contract: 201161

Project: 1940323

Date Issued: 2007/08/15

Issued to: SonoSite, Inc.

21919 30th Dr SE

Bothell, WA 98021-3904

USA

Attention: Mr. Matt Hedlund

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US'



Issued by: Cindy Johnson, ASCT.

Authorized by: Ken Rutledge, Product Group Manager

PRODUCTS

CLASS 8750 01 - MEDICAL ELECTRICAL EQUIPMENT

CLASS 8750 81 - MEDICAL ELECTRICAL EQUIPMENT - Certified to US Standards

Diagnostic Ultrasound System and accessories, Model iLook 15, iLook 25, iLook Stand, and iLook Docking Station, rated 100-240 V, 50-60 Hz, 1.2-0.6 A (with P.S. 3292 or 3481) or 2.0-1.0 A (with P. S. 3941F) with Class I Power Supply or Battery Operated. Type BF Applied Part.

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognized to perform certification to U.S. Standards.



Certificate: 1338884 (LR 115987)

Master Contract: 201161

Project: 1940323

Date Issued: 2007/08/15

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 601.1 - Medical Electrical Equipment Part 1: General Requirements for Safety

CAN/CSA-C22.2 No 60601-1-1-02 - Medical Electrical Equipment Part 1: General Requirements for Safety - 1. Collateral Standard: Safety Requirements for Medical Electrical Systems.

CAN/CSA-C22.2 No. 60601-2-37-03 -Medical Electrical Equipment - Part 2-37: Particular Requirements for the Safety of Ultrasonic Medical Diagnostic and Monitoring Equipment

CAN/CSA-C22.2 No. 60601-2-37-03, Amendment 1:2005 Medical electrical equipment - Part 2-37: Particular requirements for the safety of ultrasonic medical diagnostic and monitoring equipment

UL Std. 60601-1 (1st Edition) - Medical Electrical Equipment Part I: General Requirements for Safety

IEC 60601-1-1 (2000) - Medical Electrical Equipment Part 1: General Requirements for Safety - 1. Collateral Standard: Safety Requirements for Medical Electrical Systems.

IEC 60601-2-37 (2004-10) - Ed. 1.1 Consolidated Edition. Medical electrical equipment - Part 2-37: Particular requirements for the safety of ultrasonic medical diagnostic and monitoring equipment

IEC 60601-2-37-am1 (2004-08) Amendment 1 - Medical electrical equipment - Part 2-37: Particular requirements for the safety of ultrasonic medical diagnostic and monitoring equipment

Subject to the following qualifications:

- The equipment has not been investigated for the protection against hazards of explosions in medically used rooms.
- The main supply cord set provided with the equipment must be an approved type acceptable to the authorities in the country where the equipment is sold.
- The user replaceable mains (line) fuse must be an approved type acceptable to the authorities where the equipment is sold.
- Evaluated to IEC/CSA 601-1 Amendment 2 excluding requirements for Electromagnetic compatibility (Clause 36), Biocompatibility (Clause 48) and Programmable Electronic Systems (Clause 52.1).



Supplement to Certificate of Compliance

Certificate: 1338884

Master Contract: 201161

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Product Certification History

Project	Date	Description
1940323	2007/08/15	Update report to include alternative power supply.
1832337	2006/09/11	Correction of 60601-1-1 standard edition.
1826035	2006/08/23	Update report to include optional Printer.
1771005	2006/05/30	Update report to update standards.
1558963	2004/05/21	Update to add an alternate Power Supply.
History		
1353211	2002/09/18	Update to Report to include Model iLook 25 and iLook Stand.
1338884	2002/08/06	Original Certification.



Ref. Certif. No.

CA/8935/CSA

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE

CERTIFICAT D'ESSAI OC

Product / Produit

Diagnostic Ultrasound System and Accessories

Name and address of the applicant / Nom et adresse du demandeur

SonoSite, Inc.,
21919 - 30th Drive S.E., Bothell, WA 98024-3904 USA

Name and address of the manufacturer / Nom et adresse du fabricant

Same as applicant

Name and address of the factory / Nom et adresse de l'usine

Same as applicant

Note: When more than one factory, please report on page 2 / Note: Lorsque il y plus d'une usine, veuillez: utiliser la 2^{eme} page

Additional Information on page 2

Ratings and principal characteristics / Valeurs nominales et caractéristiques principales

100-240 V, 50-60 Hz, 1.2-0.6 A (with P.S. 3292 or 3481) or 2.0-1.0 A (with P.S. 3941F) with Class I Power Supply or Battery Operated

Trademark (if any) / Marque de fabrique (si elle existe)



Model / Type Ref. / Ref. De type

iLook 15, iLook 25, iLook Stand, and iLook Docking Station

Additional information (if necessary may also be reported on page 2) / Les informations complémentaires (si nécessaire, peuvent être indiqués sur la 2^{eme} page

Additional Information on page 2

A sample of the product was tested and found to be in conformity with / Un échantillon de ce produit a été essayé et a été considéré conforme à la

IEC 60601-1 2:1988
Amendment No 1 (1991) and Amendment No 2 (1995), excluding requirements for Electromagnetic Compatibility (Clause 36), Biocompatibility (Clause 48) and Programmable Electronic Systems (Clause 52.1); 60601-1-1(2000), 60601-2-37 (2001) and CA and US. National Differences, per CB Bulletin 112a.

As shown in the Test Report Ref. No. which forms part of this Certificate / Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat

201161-1338885 (Project 1940324)

This CB Test Certificate is issued by the National Certification Body / Ce Certificat d'essai OC est établi par l'Organisme National de Certification



CSA International
178 Rexdale Boulevard
Toronto, ON M9W 1R3

Date: August 16, 2007

Signature: Gianluca Arcari, P.Eng., MBA